

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/22/2015

mmb

Instrument Serial Number: 010574

As of 22-May-15

Start Date	End Date	Agency Name
22-May-15		DFS Central Lab
19-Dec-14	22-May-15	Chesterfield County PD
10-Dec-14	19-Dec-14	DFS Central Lab
08-Jul-14	10-Dec-14	Poquoson PD
11-Jun-14	08-Jul-14	DFS Central Lab
20-Feb-13	11-Jun-14	Hopewell PD
17-Oct-12	20-Feb-13	DFS Central Lab
06-Jan-09	17-Oct-12	Hopewell PD
30-Jun-08	06-Jan-09	DFS Central Lab

INTOX EC/IR II

Quality Assurance Worksheet

Instrument Serial Number 010574 Worksheet Start Date 5/22/2015

Location Chesterfield Adult Detention Center

Address 6900 Mimms Dr Chesterfield VA 23832

DFS Technician Michelle Billeter License No. 26417

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 757 Reference Barometer (mm HG) 760

Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.301
0.300	0.299	0.290	0.307	Sample 2	0.301
Precision		sample min	sample max	Sample 3	0.301
0		0.301	0.301		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		


Dry gas standard Lot No. (with tank no.) AG509101-44

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Notes: [Handwritten notes in cursive script, mostly illegible]

Instrument Serial Number

010574

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Ballito

Date

5/22/15

Issuing Analyst

Michelle Ballito

Date

5/22/15

JPD

mmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETER, MICHELLE, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/22/2015
TEST EQUIPMENT NUMBER 010574		

RESULTS: TIME SAMPLE TAKEN 08:16 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 22 May 2015 at 11:17

mmB

Test Results

Instrument Serial Number 010574

Test # 001028 Subject Test

Test Location 1 Department of
Test Date 22 May 2015

Test Location 2 Forensic Science
Test Time 08:10
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

Card Serial Number 126417

License Number 26417

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 22 May 2015 End Time 08:18

Result Time 08:16

Result Date 22 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 08:10

Data Type BLK

Sample Value 0.000

Sample Time 08:11

Data Type CHK

Sample Value 0.100

Sample Time 08:12

Data Type BLK

Sample Value 0.000

Sample Time 08:13

Data Type SUBJ

Sample Value 0.000

Sample Time 08:14

Data Type BLK

Sample Value 0.000

Sample Time 08:15

Data Type SUBJ

Sample Value 0.000

Sample Time 08:16

Data Type BLK

Sample Value 0.000

Sample Time 08:17

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG509101-44

Standard Expiration Date 04/01/2017

Tank Pressure 973

Barometric Pressure 757 mmHg

Blow Sample Number 1 Blow Duration 3.91 sec

Blow Volume 1764 cc End-of-Blow Time 08:14

Blow Sample Number 2 Blow Duration 4.08 sec

Blow Volume 1800 cc End-of-Blow Time 08:16

Tamper Evident Stamp 2936499e

Test Status Code 0

Test Status Success

mmB

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010574 Test Number: 1029

Test Date: 05/22/2015 Test Time: 08:19 EDT

Dry Gas Target: 0.100

Lot Number: AG509101-44 Exp Date: 04/01/2017

Tank Pressure: 968 psi Barometric Pressure: 757 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	08:20
CHK	0.100	08:20
BLK	0.000	08:22
CHK	0.100	08:23
BLK	0.000	08:24
CHK	0.100	08:25

Test Status: Success

Calibration CRC: C5BBBC31

MmB

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010574 Test Number: 1030

Test Date: 05/22/2015 Test Time: 08:36 EDT

Dry Gas Target: 0.299

Lot Number: AG506501-06 Exp Date: 03/06/2017

Tank Pressure: 517 psi Barometric Pressure: 757 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	08:37
CHK	0.301	08:38
BLK	0.000	08:39
CHK	0.301	08:40
BLK	0.000	08:42
CHK	0.301	08:43

Test Status: *Success*

Calibration CRC: C5BBBC31